

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 3/13/2015

Handwritten initials and date:
JMS
2/17

Instrument Serial Number: 010469	As of 13-Mar-15
---	------------------------

Start Date	End Date	Agency Name
12-Mar-15		DFS Central Lab
04-Nov-14	12-Mar-15	Russell County SO
23-Oct-14	04-Nov-14	DFS Central Lab
20-May-14	23-Oct-14	Middlesex County Sheriff's Office
07-May-14	20-May-14	DFS Central Lab
10-Oct-12	07-May-14	Rockingham County SO
27-Sep-12	10-Oct-12	DFS Central Lab
31-May-11	27-Sep-12	Farmville PD
09-Oct-09	31-May-11	DFS Central Lab
16-Oct-08	09-Oct-09	Chesterfield County PD
08-May-08	16-Oct-08	DFS Central Lab

INTOX EC/IR II
Quality Assurance Worksheet

KNIS
JPP

Instrument Serial Number 010469 Worksheet Start Date 3/12/2015
Location Russell County
Address 42 Court Ave., Lebanon, VA 24266
DFS Technician Heather Stanton License No. 18910
☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 715 Reference Barometer (mm HG) 717
Reference Barometer(RB)Serial # 009113 RB Calibration Due 4/28/2015

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	
0.300	0.282	0.273	0.290	0.293	
Precision		sample min	sample max	0.293	
0		0.293	0.293	0.293	

Standard (sea level)	PA Target	minimum	maximum	Sample 1	
0.100	0.094	0.091	0.097	0.096	
Precision		sample min	sample max	0.096	
0		0.096	0.096	0.096	


WARNING

Dry gas standard Lot No. (with tank no.) AG404902-07
☐ Replaced dry gas standard (+O-ring)
☐ Installed at Location
☒ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:



Instrument Serial Number

010469

Certification Date

☐ Calibrated☐ Certified☐ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Heath W. Stach

Date

3/13/15

Issuing Analyst

Jenne Dufy

Date

3/16/15




COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

HMS
JPD

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 18910	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 03/12/2015
TEST EQUIPMENT NUMBER 010469		

RESULTS: TIME SAMPLE TAKEN 08:38 EDT
0.00
SAMPLE'S ALCOHOL CONTENT _____ GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20 _____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 12 Mar 2015 at 09:19

Handwritten initials: HNS and JPD

Test Results

Instrument Serial Number 010469

Test # 002462 Subject Test

Test Location 1 Department of
Test Date 12 Mar 2015

Test Location 2 Forensic Science
Test Time 08:31
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name STANTON
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial IV

Card Serial Number 118910

License Number 18910

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 12 Mar 2015 End Time 08:39

Result Time 08:38

Result Date 12 Mar 2015

Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 08:32

Data Type BLK

Sample Value 0.000

Sample Time 08:33

Data Type CHK

Sample Value 0.096

Sample Time 08:33

Data Type BLK

Sample Value 0.000

Sample Time 08:35

Data Type SUBJ

Sample Value 0.000

Sample Time 08:35

Data Type BLK

Sample Value 0.000

Sample Time 08:36

Data Type SUBJ

Sample Value 0.000

Sample Time 08:38

Data Type BLK

Sample Value 0.000

Sample Time 08:39

Standard Type Dry Gas Std

Standard Value 0.094

Standard Lot Number AG404902-07

Standard Expiration Date 02/18/2016

Tank Pressure 835

Barometric Pressure 715 mmHg

Blow Sample Number 1 Blow Duration 3.25 sec

Blow Volume 1763 cc End-of-Blow Time 08:35

Blow Sample Number 2 Blow Duration 3.42 sec

Blow Volume 1787 cc End-of-Blow Time 08:38

Tamper Evident Stamp 5b361050

Test Status Code 0

Test Status Success

AMS
JPD

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010469 Test Number: 2463

Test Date: 03/12/2015 Test Time: 08:41 EDT

Dry Gas Target: 0.094

Lot Number: AG404902-07 Exp Date: 02/18/2016

Tank Pressure: 830 psi Barometric Pressure: 715 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	08:41
CHK	0.096	08:42
BLK	0.000	08:44
CHK	0.096	08:44
BLK	0.000	08:46
CHK	0.096	08:46

Test Status: *Success*

Calibration CRC: B8F2C78C

AMS
JAD

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010469 Test Number: 2464

Test Date: 03/12/2015 Test Time: 08:49 EDT

Dry Gas Target: 0.282

Lot Number: AG428003-02 Exp Date: 10/07/2016

Tank Pressure: 487 psi Barometric Pressure: 715 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	08:50
CHK	0.293	08:51
BLK	0.000	08:53
CHK	0.293	08:54
BLK	0.000	08:55
CHK	0.293	08:56

Test Status: *Success*

Calibration CRC: B8F2C78C

Out of tolerance for accuracy per QA sheet.
AMS
3/12/15

AMS
JRP

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010469 Test Number: 2465

Test Date: 03/12/2015 Test Time: 08:59 EDT

Dry Gas Target: 0.282

Lot Number: AG428003-04 Exp Date: 10/07/2016

Tank Pressure: 110 psi Barometric Pressure: 715 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	08:59
CHK	0.293	09:00
BLK	0.000	09:02
CHK	0.293	09:03
BLK	0.000	09:05
CHK	0.293	09:06

Test Status: *Success*

Calibration CRC: B8F2C78C

Out of tolerance for accuracy per QA Sheet.
AMS
3/12/15